

---

## PEM-ProLib++

### Programming Library for Plasma Emission Measurement


---

For developing your own plasma emission measurement applications using our TranSpec spectrometers, we provide our powerful and very easy-to-use programming library **PEM-ProLib++**.

With PEM-ProLib++ the entire spectra data acquisition, like scanning the diode array, raw data averaging, possible dark current correction and the emission spectra normalization is fully encapsulated in just a few simple function calls. PEM-ProLib+ gives you full access to all measured spectra (including the raw data) and easily allow to report emission trace values to an external analog out module.

- Runtime licensed Dynamic Link Library (DLL) providing standard C calls  
Compatible with common C/C++ compiler, Visual Basic and VBA (Excel), LabVIEW
- Extensive parameter checking and measurement status verification during runtime  
You hardly can do anything wrong when working with PEM-ProLib++
- Supports special external I/O module with 8-channel TTL and 4-channel analog out
- Detailed user manual as PDF file and optionally as printed document
- Demo software as Windows console application, including C/C++ source code
- See next page for a short programming example!

Technical specifications on next page ►



---

## PEM-ProLib++ Programming Library • Technical Specifications

January 2026, related to version 3.0, without guarantee, subject to changes.

### Minimum Hardware and Software Requirements

- Standard PC/Laptop with Windows 10 or Windows 11
- C/C++ development system (MS Visual Studio recommended), Delphi, Visual Basic or VBA (Excel), LabVIEW
- TranSpec Process Spectrometer
- PEM-ProVis Professional is recommended, but not required

### Programming Example

The following short programming example may demonstrate how easy to use the PEM-ProLib++ is. As an example, we code the fully automated measurement of a 10 times averaged emission spectrum with subsequent analog output of an emission trace value:

```
// Step 1: open and initialize spectrometer
PEMPROLIB_SPECHARDWARE sSpecHardwareInfo;
PEMProLib_OpenSpectrometer( PEMPROLIB_TRANSPEC_19Z, &sSpecHardwareInfo );

// Step 2: setup measurement parameter:
PEMPROLIB_MEASPARA sMeasPara;
sMeasPara.dIntegrationTime = 20.0;           // 20 ms integration time
sMeasPara.bEnableAverage = 1;              // averaging on
sMeasPara.iNumberAverage = 10;            // 10 scans for averaging
PEMProLib_SetMeasPara( &sMeasPara );       // notify settings to spectrometer

// Step 3: open and initialize external digital/analog module
PEMProLib_USB3110_OpenDevice( PEMPROLIB_DIGITAL-8OUT0IN, PEMPROLIB_ANALOG_UNIPOLAR );

// Step 4: perform measurement of an averaged emission spectrum
PEMProLib_RunMeasSpectrum();               // start measurement

PEMPROLIB_SPECSTATUS sSpecStatus;
PEMProLib_GetSpecStatus( &sSpecStatus );   // wait until measurement is done
while ( sSpecStatus.bRunSpectrum )
    PEMProLib_GetSpecStatus( &sSpecStatus );

// Step 5: retrieve measured emission spectrum
PEMPROLIB_SPECDATA sSpecData;
PEMProLib_GetSpectrumData( PEMPROLIB_SPECTRUM_EMISSION, &sSpecData );

// Step 6: compute and report emission trace value at 254 nm, for instance:
double dTraceValue;
PEMProLib_GetTraceValue( 254.0 , &dTraceValue );

// normalize trace value to +10 V and report at analog channel 0
....
PEMProLib_USB3110_SetAnalogOut( 0 , dTraceValueAsVolt );
```

---

**Note** TranSpec is a registered German trademark of Dipl.-Ing. (FH) Th. Fuchs, Engineer's Office for Applied Spectroscopy. All other product names mentioned are or possibly might be trademarks or registered trademarks of their owners.

Revision Jan-2026